



Human Leptin ELISA

Catalog Number EA-0202

(For Research Use Only)

Introduction

Human leptin is a protein of 167 amino acids that plays a key role in regulating energy intake and energy expenditure, including the regulation (decrease) of appetite and (increase) of metabolism. Leptin is produced by adipose tissue and the level of circulating leptin is directly proportional to the total amount of fat in the body. Once leptin has bound to the Ob-Rb receptor, it activates Stat3, which is phosphorylated and travels to the nucleus where it mediates gene expression. One of the main effects on gene expression is the down-regulation of the expression of endocannabinoids, responsible—among their many other functions—for increasing appetite. There are other intracellular pathways activated by leptin, but less is known about how they function in this system. Although leptin is a circulating signal that reduces appetite, in general, obese people who are resistant to the effects of leptin, in the same way that people with type 2 diabetes are resistant to the effects of insulin (1). Leptin is also found to stimulate endothelial cell proliferation and angiogenesis.

Principle of the assay

Leptin ELISA is based on the principle of a solid phase enzyme-linked immunosorbent assay. The assay utilizes a monoclonal antibody against a distinct determinant on human leptin for immobilization on the microtiter wells and a biotinated mouse anti-leptin antibody along with streptavidin conjugated to horseradish peroxidase (HRP) for detection. The test sample is allowed to react simultaneously with the two antibodies, resulting in the leptin molecules being sandwiched between the solid phase and enzyme-linked antibodies. After incubation, the wells are washed to remove unbound-labeled antibodies. A HRP substrate, TMB, is added to result in the development of a blue color. The color development is then stopped with the addition of stop solution changing the color to yellow. The concentration of leptin is directly proportional to the color intensity of the test sample. Absorbance is measured spectrophotometrically at 450 nm.



Diagram of ELISA

Materials provided with the kit

- 96 well microplate coated with a mouse anti-human leptin antibody (4°C).
- Biotin labeled mouse anti-human leptin antibody (-20°C).
- Streptavidin-HRP conjugate (4°C).
- Recombinant human leptin standard (500ng/ml) (-20°C).
- 1X Diluent buffer (4°C).
- 5X Assay wash buffer (RT)
- Substrate (4°C).
- Stop Solution (4°C).

Material required but not provided


- Microplate reader capable of measuring absorbance at 450 nm
- Deionized or distilled water.

Reagent preparation before starting experiment

- Dilute the 5x Assay wash buffer to 1x buffer
40ml 5x Assay wash buffer
160ml ddH₂O
- Dilute 50 times of human recombinant leptin (500ng/ml) with 1X Diluent buffer to 10ng/ml and then 2-fold serial dilutions. Add 4ul Human Recombinant Leptin in 200ul 1X Diluent Buffer (See Step 2 in "Assay Procedure" for detailed instruction)
- Use serum-free conditioned media or original or 10-fold diluted sera. Sera can be diluted with 1 X Diluent buffer. When serum-containing conditioned media is required, be sure to use serum as a control.
- Dilute 400 times of biotin labeled mouse anti-human leptin antibody with 1X Diluent buffer before use.
- Dilute 200 times of streptavidin-HRP with 1X Diluent buffer before use.

Assay procedure

1. Cut the sealing film over the plate and remove it from the desired number of wells. Make sure the rest of wells are well sealed.
2. See instruction and diagram below for standard preparation.



- a. Add 200ul 1X Diluent buffer to the 1st well. Add 100ul 1X Diluent Buffer to the rest wells of strip.
- b. Add appropriate amount of protein recombinant (follow instruction in "Reagent Preparation")
- c. Mix dilutions in 1st well and transfer 100ul from the 1st well to the next dilution. (See picture) Incubate each well for 1 hr at room temperature with gentle shaking

3. Add 100ul of sample per well and incubate for 1 hour at room temperature with gentle shaking.
4. Aspirate each well and wash by adding 200ul of 1X Assay wash buffer. Repeat the process three times for a total of three washes. Complete removal of liquid at each wash. After the last wash, remove any remaining liquid by inverting the plate against clean paper towels.
5. Add 100ul of diluted biotin-labeled mouse anti-human Leptin antibody to each well and incubate for 1 hour at room temperature with gentle shaking.
6. Repeat the aspiration/wash as in step 4.

7. Add 100 ul of diluted streptavidin-HRP conjugate to each well and incubate for 45 min at room temperature with gentle shaking.
8. Repeat the aspiration/wash as in step 4.
9. Add 100ul of substrate to each well and incubate for 5-10 minutes.
10. Add 50ul of Stop solution to each well. The color in the wells should change from blue to yellow.
11. Determine the optical density of each well with a microplate reader at 450 nm within 30 minutes.

References

1. Considine RV, Sinha MK, Heiman ML, Kriauciunas A, Stephens TW, Nyce MR, Ohannesian JP, Marco CC, McKee LJ & Bauer TL (1996). "Serum Immunoreactive-Leptin Concentrations in Normal-Weight and Obese Humans". *N Engl J Med* **334** (5): 292-295. PMID 8532024.

Example of standard curve

