



Human IL-12 ELISA

Catalog Number EA-0510

(For Research Use Only)

Introduction

Interleukin 12 (IL-12) is an interleukin that plays important roles in the activities of natural killer cells and T lymphocytes, the differentiation of naive T cells into Th1 cells, the growth and function of T cells, the production of interferon-gamma (IFN- γ) and tumor necrosis factor-alpha (TNF- α) from T and natural killer (NK) cells, and IL-4 mediated suppression of IFN- γ . It also plays a role in anti-angiogenic activity by increasing production of IFN- γ , which in turn increases the production of IP-10. IL-12 is naturally produced by dendritic cells, macrophages and human B-lymphoblastoid cells in response to antigenic stimulation.

Principle of the assay

IL-12 ELISA is based on the principle of a solid phase enzyme-linked immunosorbent assay. The assay utilizes goat anti-human IL-12 for immobilization on the microtiter wells and biotinylated goat anti-human IL-12 antibodies along with streptavidin conjugated to horseradish peroxidase (HRP) for detection. The test sample is allowed to react simultaneously with the two antibodies, resulting in the IL-12 molecules being sandwiched between the solid phase and enzyme-linked antibodies. After incubation, the wells are washed to remove unbound-labeled antibodies. A HRP substrate, TMB, is added to result in the development of a blue color. The color development is then stopped with the addition of Stop Solution changing the color to yellow. The concentration of IL-12 is directly proportional to the color intensity of the test sample. Absorbance is measured spectrophotometrically at 450 nm.




Diagram of ELISA

Materials provided with the kit

- 96 well microplate coated with goat anti-human IL-12 antibodies (4°C).
- Biotin labeled goat anti-human IL-12 antibodies (-20°C).
- Streptavidin-HRP conjugate (4°C).
- Recombinant human IL-12 standard (1000ng/ml) (-20°C).
- 1X Diluent buffer (4°C).
- 5X Assay wash buffer (RT)
- Substrate (4°C).
- Stop Solution (4°C).

Material required but not provided

- Microplate reader capable of measuring absorbance at 450 nm
- Deionized or distilled water.

Reagent preparation before starting experiment

- Dilute the 5x Assay wash buffer to 1x buffer
40ml 5x Assay wash buffer
160ml ddH₂O
- Dilute 500 times of human recombinant IL-12 (1000ng/ml) with 1X Diluent buffer to 2000pg/ml and then 2-fold serial dilutions.
- Dilute 400 times of biotin labeled goat anti-human IL-12 antibodies with 1X Diluent buffer before use.
- Dilute 200 times of streptavidin-HRP with 1X Diluent buffer before use.

Assay procedure

1. Cut the sealing film over the plate and remove it from the desired number of wells. Make sure the rest of wells are well sealed.
2. Add 100 µl of Standard, control, or sample per well and incubate for 1 hour at room temperature with gentle shaking.
3. Aspirate each well and wash by adding 200µl of 1X Assay wash buffer. Repeat the process three times for a total of three washes. Complete removal of liquid at each wash. After the last wash, remove any remaining liquid by inverting the plate against clean paper towels.
4. Add 100 µl of diluted biotin-labeled goat anti-human IL-12 antibodies to each well and incubate for 1 hour at room temperature with gentle shaking.
5. Repeat the aspiration/wash as in step 3.
6. Add 100 µl of diluted streptavidin-HRP conjugate to each well and incubate for 45 min at room temperature with gentle shaking.
7. Repeat the aspiration/wash as in step 3.
8. Add 100µl of substrate to each well and incubate for 5-30 minutes.
9. Add 50µl of Stop solution to each well. The color in the wells should change from blue to yellow.
10. Determine the optical density of each well with a microplate reader at 450 nm within 30 minutes.

Example of standard curve

